



How to update the status webpage to the latest version

This update will correct the Java certificate issue with Java Rev 45

Models: VM505E, VM510E

Download and unzip StatusPageUpdate.zip. A folder Status Page Update will be created. Open the folder.

Double click on **updatevm500e.bat**. Enter the IP address of the device to be updated and then press the enter key.

The status pages will be automatically updated.

Exit the command prompt.

To test, open a browser and type in the IP address.

Verify that the status page is version 4.1.9 and the Daily Report is 4.1.8. Both will have the new blue background.

Unit ID: ICRC FF 88 (LOANER) Model: VM505E
Last Updated: 11/07/13 10:48 AM

Buttons: Refresh Status, Download Log, Show Log, Graph Time Frame, Graph Log, Print Graph, Daily Report, Trending Time Frame, Enable Trending, Rev. 4.1.9, Help

Select	Sensor	Name	Reading	Units	Status	Lower Limit	Upper Limit	Time Out Of Limit
<input type="checkbox"/>	Sensor 1	ICRC FR 14	-22.5	*C	OK	-87.0	-65.0	28
<input type="checkbox"/>	Sensor 2	ICRC FR 15 - Off	-22.5	*C	OK	-87.0	30.0	0
<input type="checkbox"/>	Sensor 3	ICRC FR 19	-6.6	*C	OK	-87.0	-65.0	28
<input type="checkbox"/>	Sensor 4	ICRC Cabrini 2	-6.5	*C	OK	-87.0	-65.0	28
<input type="checkbox"/>	Sensor 5	ICRC FR 21	-6.8	*C	OK	-87.0	-65.0	28
<input type="checkbox"/>	Sensor 6	ICRC FR 33		*C	N/C	-87.0	-65.0	0
<input type="checkbox"/>	Sensor 7	ICRC FR 34		*C	N/C	-87.0	-65.0	0
<input type="checkbox"/>	Sensor 8	ICRC FR 20		*C	N/C	-87.0	-65.0	0

Temperature Guard Print Log - Google Chrome
192.168.0.81/index_logprint.html

Unit ID1: ICRC FF 88 (LOANER) Model: VM505E
4 Hours
Go
Rev. 4.1.8 Help