

FUNCTION TESTED	RANGE	APPLIED VALUE	AS FOUND VALUE	TOTAL UNCERTAINTIES	TOLERANCE
INPUT 7					
TEMPERTURE	-149 - 78	-70.0°	<u>-69.6</u>	<u>±(0.07+0.06) °F</u>	< 1.0°F
ACCURACY	°F	-15.0°	<u>- 14.6</u>	<u>±(0.09+0.06) °F</u>	< 1.0°F
RTD 1000Ω		36.0°	<u>36.2</u>	<u>±(0.09+0.06) °F</u>	< 1.0°F
4W - 2W		46.0°	<u>46.3</u>	<u>±(0.09+0.06) °F</u>	< 1.0°F
		78.0°	<u>77.7</u>	<u>±(0.09+0.06) °F</u>	< 1.0°F
INPUT 8					
TEMPERTURE	-149 - 78	-70.0°	<u>-69.5</u>	<u>±(0.07+0.06) °F</u>	< 1.0°F
ACCURACY	°F	-15.0°	<u>- 14.5</u>	<u>±(0.09+0.06) °F</u>	< 1.0°F
RTD 1000Ω		36.0°	<u>36.3</u>	<u>±(0.09+0.06) °F</u>	< 1.0°F
4W - 2W		46.0°	<u>46.3</u>	<u>±(0.09+0.06) °F</u>	< 1.0°F
		78.0°	<u>77.6</u>	<u>±(0.09+0.06) °F</u>	< 1.0°F

Equipment Used 1237 _____

Condition as Found: In tolerance X Out of tolerance _____ Op Fail _____
 Condition as Left : In tolerance X Out of tolerance _____

Comments:

Technician: CR Pate

CAL TIME 1.0
 REP TIME _____